

SOLUTION CHANGE

 SHERRILL CITY
 POLICE DEPT
 DATAMASTER dmt: 115108

Date: 05/01/2020

OPERATOR NAME:
 ROBERT J DRAKE
 Permit #: 27298
 Operator Agency/Dept: SHERRILL CITY PD
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:42
INTERNAL STANDARD	VERIFIED	07:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:45
BLANK TEST	0.00	07:46

