

SUPERVISOR TEST

 SHERRILL CITY
 POLICE DEPT
 DATAMASTER dmt: 115108

Date: 11/27/2019

OPERATOR NAME:
 ROBERT J DRAKE
 Permit #: 27298
 Operator Agency/Dept: SHERRILL CITY PD
 Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:56
INTERNAL STANDARD	VERIFIED	06:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:58
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	07:00

