

SUPERVISOR TEST

-----  
SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108  
-----

Date: 01/15/2020

OPERATOR NAME:  
SHAWN R PLATT  
Permit #: 30289  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:50
INTERNAL STANDARD	VERIFIED	06:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53

