

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 06/03/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:17
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21

